

ARM FY2025 Aerosol Operations Plan

O Mayol-Bracero	J Uin
S Smith	J Shilling
M Zawadowicz	A Singh
J Creamean	A Sedlacek
C Hayes	D Campos DeOliveira
M Salatti	O Enekwizu
C Flynn	M Petters
S Petters	A Theisen

May 2025



DISCLAIMER

This report was prepared as an account of work sponsored by the U.S. Government. Neither the United States nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the U.S. Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the U.S. Government or any agency thereof.

ARM FY2025 Aerosol Operations Plan

O Mayol-Bracero, Brookhaven National Laboratory (BNL)

J Uin, BNL

S Smith, BNL

J Shilling, Pacific Northwest National Laboratory

M Zawadowicz, BNL

A Singh, BNL

J Creamean, Colorado State University

A Sedlacek, BNL

C Hayes, BNL

D Campos DeOliveira, BNL

M Salatti, BNL

O Enekwizu, BNL

C Flynn, University of Oklahoma

M Petters, University of California, Riverside (UCR)

S Petters, UCR

A Theisen, Argonne National Laboratory

May 2025

Mayol-Bracero, O, J Uin, S Smith, J Shilling, M Zawadowicz, A Singh, J Creamean, A Sedlacek, C Hayes, D Campos DeOliveira, M Salatti, O Enekwizu, C Flynn, M Petters, S. Petters, and A Theisen. 2025. ARM FY2025 Aerosol Operations Plan. U.S. Department of Energy, Atmospheric Radiation Measurement user facility, Richland, Washington. DOE/SC-ARM-TR-311.

Work supported by the U.S. Department of Energy,
Office of Science, Office of Biological and Environmental Research

Acronyms and Abbreviations

3D	three-dimensional
AAF	ARM Aerial Facility
ACSM	aerosol chemical speciation monitor
ACSM-CDCE	ACSM, Corrected for Composition-Dependent Collection Efficiency Value-Added Product
ACTRIS	Aerosol, Clouds and Trace Gases Research Infrastructure
AETH	aethalometer
AFC	ARM field campaign
AFS	aerosol flux measurement system
AMF	ARM Mobile Facility
AMICE-2	Absorption Measurements InterComparison Experiment 2
AMSG	Aerosol Measurement Science Group
AOD	aerosol optical depth
AODBE	Aerosol Optical Depth Best Estimate Value-Added Product
AOP	Aerosol Operations Plan
AOS	Aerosol Observing System
AOSMET	Aerosol Observing System meteorological instruments
APS	aerodynamic particle sizer
ARM	Atmospheric Radiation Measurement
AWARE	ARM West Antarctic Radiation Experiment
BNF	Bankhead National Forest
BNL	Brookhaven National Laboratory
CAMS	Center for Aerosol Measurement Science
CAPE-k	Cape kennaook/Grim, Tasmania
CAPS	cavity attenuated phase shift monitor
CAPS-SSA	cavity attenuated phase shift monitor-single-scattering albedo
CARGO-ACT	Cooperation and AgReements enhancing Global interOperability for Aerosol, Cloud, and Trace gas research infrastructures
CCN	cloud condensation nuclei
CHARMS	Cryogenic, High-Accuracy Refraction Measuring System
CLAP	continuous light absorption photometer
CoURAGE	Coast-Urban-Rural Atmospheric Gradient Experiment
CPC	condensation particle counter
CPCF	condensation particle counter, fine
CPCUF	condensation particle counter, ultrafine
CRG	ARM site code for CoURAGE campaign

CSU	Colorado State University
DAQ	data acquisition system
ENA	Eastern North Atlantic
ECAPE	Eastern Pacific Cloud Aerosol Precipitation Experiment
FAIR	Findability, Accessibility, Interoperability, and Reusability
FY	fiscal year
GAW	Global Atmosphere Watch
GHG	greenhouse gases monitor
HSRL	high-spectral-resolution lidar
HTDMA	humidified tandem differential mobility analyzer
INP	ice nucleating particle, filters for ice nucleation particles
INS	ice nucleating spectrometer
IOP	intensive operational period
KCG	ARM site code for kennaook/Cape Grim campaign
MCPC	mixing condensational particle counter
ML	machine learning
NANOSMPS	nano scanning mobility particle sizer
NASA	National Aeronautics and Space Administration
NEPHDRY	nephelometer, ambient
NEPHWET	nephelometer, RH scanned
NOAA	National Oceanic and Atmospheric Administration
NOX	nitrogen oxide monitor
NSA	North Slope of Alaska
OPC	optical particle counter
OZONE	ozone monitor
PI	principal investigator
POPS	portable optical particle spectrometer
PSAP	particle soot absorption photometer
PTAAM	photothermal aerosol absorption monitor
PTI	photothermal interferometer
PTRH	pressure, temperature, and relative humidity
PTR-MS	proton transfer reaction-mass spectrometer
QA	quality assurance
QC	quality control
RH	relative humidity
RL	Raman lidar
RLPROF-FEX	Raman Lidar Vertical Profiles-Feature Detection and Extinction Value-Added Product

SAIL	Surface Atmosphere Integrated Field Laboratory
SGP	Southern Great Plains
SMPS	scanning mobility particle sizer
SP2	single-particle soot photometer
SP2-XR	extended-range SP2
TBD	to be determined
TBS	tethered balloon system
TOF-ACSM	aerosol chemical speciation monitor–time of flight
TRACER	Tracking Aerosol Convection Interactions Experiment
TROPOS	Leibniz Institute for Tropospheric Research
UHSAS	ultra-high-sensitivity aerosol spectrometer
VAP	value-added product
WCCAP	World Calibration Center for Aerosol Physics
WMO GAW	World Meteorological Organization Global Atmosphere Watch
WP	work package

Contents

Acronyms and Abbreviations	iii
1.0 Introduction	1
2.0 AOS Instrumentation.....	1
2.1 Instrument Tiers	3
2.1.1 Standard AOS.....	3
2.1.2 Additional AOS Instruments.....	3
2.1.3 Aerosol Nodes	3
3.0 Outcomes from the 2024 Aerosol Operations Plan	4
4.0 ARM AOS Instrument Operations	6
4.1 FY25 Operations	6
4.1.1 Aerosol Observing System.....	6
4.1.2 Hygroscopicity Plan	6
4.1.3 Data Products	7
4.1.4 Ice Nucleating Particles.....	7
4.2 Calibration.....	7
4.2.1 Calibration Schedule	7
4.2.2 Intensive Operational Periods	8
4.2.3 Improving Calibration Processes.....	9
4.2.4 Training Site Operations for Calibrations	10
4.3 FY25 Engineering and Development	10
4.3.1 Aerosol Node Development	11
4.3.2 Distributed Aerosol Size Distribution Network for CoURAGE	12
4.3.3 Aerosol Flux Measurement System (AFS) Development	12
4.3.4 Center for Aerosol Measurement Science (CAMS) and Setting up SMPS and CPC Gold-Standard Calibration Capabilities for ARM AOS Instruments	13
4.3.5 CARGO-ACT.....	13
4.3.6 Absorption Measurements InterComparison Experiment 2 (AMICE-2)	14
4.3.7 Distributed Aerosol Measurement for Urban Deployments.....	14
4.3.8 New Trace Gas Analyzer	14
4.3.9 CoURAGE NO _x Measurements	14
5.0 ARM AOS Data Products.....	15
5.1 FY24 Accomplishments and Ongoing Activities.....	15
5.2 FY25 Planned Activities	15
6.0 References	16
Appendix A – 2024 Aerosol Operations Plan Completed Tasks.....	A.1
Appendix B – 2024 Aerosol Measurement In-Progress Tasks	B.1
Appendix C – Download Metrics 1/1/2018-7/7/2023.....	C.1

Tables

1	Matrix of aerosol instruments and the observatories where they are deployed.....	2
2	Standard AOS instruments.	3
3	Additional AOS instruments.	3
4	Outcomes from FY24 Aerosol Operations Plan (Theisen et al 2024). Green-shaded tasks are completed and tan-shaded are in progress.....	5
5	Planned calibration activities for each of the ARM observatories.	8
6	Calibration frequency and needs for each instrument type.	8
7	CAMS calibration activities for FY25.	10
8	FY25 planned aerosol engineering and development activities.	11
9	Proposed FY25 milestones and progress so far.....	12
10	Revised milestones for the AFS in 2025.	13
11	Instrument download metrics, 2018-2023.	C.1

1.0 Introduction

The U.S. Department of Energy's Atmospheric Radiation Measurement user facility (ARM) deploys at each of ARM's observatories a suite of aerosol and trace gas (further mention of aerosols will assume inclusion of trace gases) instrumentation that constitute the Aerosol Observing Systems (AOS; Uin et al. 2019). ARM currently deploys five AOSs, one each at:

- Southern Great Plains (SGP)
- Eastern North Atlantic (ENA)
- ARM Mobile Facilities (AMF1, 2, and 3).

Aerosol measurements at ARM's North Slope of Alaska (NSA) observatory have historically been made by the National Oceanic and Atmospheric Administration (NOAA) and provided to ARM through a collaboration. To complement these measurements, starting in financial year 2025 (FY25), ARM will monitor chemical composition and particle size by adding an aerosol chemical speciation monitor (ACSM), extended-range single-particle soot photometer (SP2-XR), and aerodynamic particle sizer (APS) at the NOAA facility in NSA.

As ARM's support for aerosol instrumentation increases, it is important for ARM to communicate plans and priorities for aerosol measurements to the community to maximize the benefit and planning around scientific activities and to advance confidence in ARM's aerosol measurements.

ARM is developing a yearly aerosol operations plan for every fiscal year (October 1-September 30). This started in FY24. The plan will be open and available through the ARM aerosol instrument webpages and will include:

- Review of the previous activities since the last plan
- Planned activities and their priorities for the upcoming FY
- Calibration timing and efforts
- Planned activities for data products and value-added products (VAPs).

Questions about the plan should be sent to Olga Mayol-Bracero through the mentor contact page: <https://arm.gov/connect-with-arm/organization/instrument-mentors/list#aos>.

2.0 AOS Instrumentation

Each AOS has a common set of standard instruments with additional instruments deployed as needed to ensure the best measurements at each site. The measurements include aerosol particle number concentration, size distribution, chemical composition, radiative and optical properties, hygroscopicity, concentration of trace gases, and supporting meteorological conditions. Table 1 lists instrumentation deployed at each ARM site and other instrumentation available for field campaigns such as during intensive operational periods (IOPs).

Table 1. Matrix of aerosol instruments and the observatories where they are deployed.

Instrument	AMF1	AMF2*	AMF3	SGP	ENA	NSA	IOP
Aerosol chemical speciation monitor-quadrupole (ACSM)							
Aerosol chemical speciation monitor-time of flight (TOF-ACSM)							
Aethalometer (AETH)						NOAA	
Aerodynamic particle sizer (APS)		CAPE-k					
Cavity attenuated phase shift monitor (CAPS)							
Carbon monoxide/nitrous oxide/water vapor (CO)		Y					
Continuous light absorption photometer (CLAP)						NOAA	
Cloud condensation nuclei (CCN)		Y					
Condensation particle counter (CPC, CPCF)		Y				NOAA	
Ultrafine condensation particle counter (CPCUF, CPCU)		Y					
Humidified tandem differential mobility analyzer (HTDMA)**							
Impactor (1-10 μm)		Y				NOAA	
Nano scanning mobility particle sizer (NANOSMPS)							
Nephelometer, ambient (NEPHDRY)		Y				NOAA	
Ozone (OZONE)		Y					
Scanning mobility particle sizer (SMPS)		Y					
Sulfur dioxide (SO ₂)							
Single-particle soot photometer (SP2)	IOP	CAPE-k	XR			XR	X
Ultra-high-sensitivity aerosol spectrometer (UHSAS)		CAPE-k					
Meteorological information (AOSMET)							
Filters for ice nucleating particles (INS/INP)						X	X
Nitrogen oxide 3-channel [NO, NO ₂ , NO _y] (NOX)							X
Greenhouse gases [CO ₂ , CH ₄] (GHG)							X
Y = part of the system							
Additions in FY25							
Removed from operations							

*The AMF2 AOS, except for the APS, SP2, and UHSAS, is not deployed in the current AMF2 deployment at [CAPE-k](#) (Cloud, Aerosol, and Precipitation Experiment at kennaook/Cape Grim, April 15, 2024-Sept 15, 2025) due to the overlap with the existing measurements at the Cape Grim, Tasmania aerosol stations.

** For FY25, ARM will operate a single HTDMA at AMF1 (Coast-Urban-Rural Atmospheric Gradient Experiment [CoURAGE-CRG]). After the completion of CRG, ARM will retire all HTDMA instruments.

2.1 Instrument Tiers

ARM operates a high-quality, research-grade AOS system and since 2024 started to invest in other tiers of aerosol measurements including aerosol nodes. Below are the definitions of the different instrument tiers: standard AOS, additional AOS, and aerosol nodes. Instrumentation outside the standard AOS could be operated during IOPs or field campaigns as requested.

2.1.1 Standard AOS

An ARM standard AOS system has high-quality, research-grade instrumentation. These are highly complex systems with a variety of instrumentation and sampling protocols. A standard AOS includes the following instruments, currently deployed at every AOS.

Table 2. Standard AOS instruments.

Aerodynamic particle sizer (APS)	Ozone (O ₃)
Cloud condensation nuclei counter (CCN)	Particle soot absorption photometer (PSAP)
Condensation particle counter (CPC/CPCF)	Scanning mobility particle sizer (SMPS)
Impactor	Ultra-high-sensitivity aerosol spectrometer (UHSAS)
Nephelometer (NEPH)	AOS meteorological system (AOSMET)

2.1.2 Additional AOS Instruments

Additional AOS instrumentation (Table 3) may be deployed full-time at certain sites or just during IOPs. These instruments are not the same across all AOSs and could be deployed standalone in some cases.

Table 3. Additional AOS instruments.

Aerosol chemical speciation monitor (ACSM/TOF-ACSM)	Nano scanning mobility particle sizer (NANOSMPS)
Aethalometer (AETH)	Sulfur dioxide monitor (SO ₂)
Carbon monoxide monitor (CO)	Single-particle soot photometer (SP2, SP2-XR)
Ultrafine condensation particle counter (CPCU/CPCUF)	Filters for ice nucleation particles (INS/INP)
Cavity attenuated phase shift monitor (CAPS)	

2.1.3 Aerosol Nodes

ARM is defining aerosol nodes to cover a range of possible aerosol measurement systems for deployment, from lower-cost standalone systems to research-grade instruments deployed in a small shelter. The measurements deployed will depend on the scientific priorities of the field campaigns and may include, but are not limited to, particle size distributions, particle number concentrations, and trace gases. Two

tiers of aerosol nodes are being developed within ARM, known as the mini-AOS and micro-AOS. The key differences between these two types of nodes lie in their operational complexity, measurement capabilities, and physical footprint. As noted in the Engineering Development Section, ARM is currently investing in two mini-AOS for the extended AMF3 deployment to the Bankhead National Forest (BNF) in northwest Alabama. This system will initially consist of size distribution measurements from ~ 0.01 to $10\text{ }\mu\text{m}$ and, with deployment maturity of the mini-AOS, will include additional capabilities such as optical measurements. ARM is also deploying a network of three micro-AOS using Handix portable optical particle spectrometers (POPS) during the urban AMF1 deployment, CoURAGE (Maryland, December 1, 2024–November 30, 2025).

3.0 Outcomes from the 2024 Aerosol Operations Plan

The outcomes presented here correspond to accomplishments related to the FY24 Aerosol Operations Plan (AOP; Theisen et al. 2024) and the 2018 Aerosol Measurement Plan (Mather et al. 2018). Table 4 shows completed tasks in green shading and in-progress tasks in tan shading. A short overview of the completed tasks from this plan can be found in **Appendix A** with overviews of the in-progress tasks in **Appendix B**.

Two tasks were removed from the FY24 AOP. The establishment of a cloud condensation nuclei (CCN) scan strategy was a recommendation from the 2018 Aerosol Measurement Plan. Since the CCN instrument is not built to offer the flow-scanning capability, to do this would require modification of the hardware and software. This could have an impact on manufacturer's support; therefore, this task is not part of our plan in the near future and will require further discussion with the Aerosol Measurement Science Group (AMSG). The other task that was removed was the characterization of the humidigraphs. These instruments were aging in ARM, and many were at the point of needing replacement. Instead of replacing these systems, in FY24 ARM decided to retire all humidigraphs and RH-scanned nephelometers from operations, leaving only a single nephelometer at each site.

Other changes for FY24 that impacted aerosol operations include ACSMs and humidified tandem differential mobility analyzer (HTDMAs). The Eastern North Atlantic (ENA) ACSM was removed from service and is kept as a spare. The AMF2 TOF-ACSM was installed at NSA. ARM will not operate any HTDMA instruments going forward, but will keep a single unit mothballed for potential future use. The HTDMA at CoURAGE will be kept operational until the end of the campaign. Details are presented in Section 4.1.

Table 4. Outcomes from FY24 Aerosol Operations Plan (Theisen et al 2024). Green-shaded tasks are completed and tan-shaded are in progress.

#	Task	Priority	Year Planned
1	Improve and simplify access to aerosol data	1	2018
2	Improve documentation of measurements and data streams	1	2018
3	Identify candidate data products for other communities	1	2025
4	Reduce number of HTDMAs in field	1	2018
5	Refurbishment of ENA HTDMA	2	2024
6	Removal of ACSM at ENA	3	2024
7	Retirement of humidigraphs	3	2024
8	New instrumentation for AMF3 (SO ₂ , APS, SP2-XR)	1	2024
9	Implement an inlet drying system (initially at SGP)	2	2025
10	Implement the new pressure, temperature, and relative humidity (PTRH) sensors in AOS02 and AOS03	2	2025
11	NSA additional instrument deployment	1	2024
12	Develop prioritization for ultrafine CPCs	3	2019
13	Implement second set of size distribution instruments	2	2019
14	Comparable size distribution output across instruments	1	2018
15	Absorption, scattering, and extinction closure	3	2023
16	Increase particle soot absorption photometer (PSAP) filter change at SGP	4	2023
17	New trace gas analyzer	2	2024
18	Develop plan to support detailed composition measurements	3	2019
19	Develop strategy for hygroscopic measurements	1	2023
20	Upgrade an ARM HSRL to support aerosol retrievals	1	2024
21	Support lidar/radiometer retrieval development	1	2025

4.0 ARM AOS Instrument Operations

4.1 FY25 Operations

4.1.1 Aerosol Observing System

ARM will continue operating AOSs at SGP, ENA, AMF1 (CoURAGE), and AMF3 (BNF) in FY25. Additional instrumentation has been added to the NOAA NSA facility at the end of FY24 (see Table 1 for the specific instruments). During the Cape kennaook/Grim, Tasmania campaign (CAPE-k) ending in September 2025, only a small subset of instrumentation (APS, SP2, and ultra-high-sensitivity aerosol spectrometer [UHSAS]) is being deployed at the AMF2 as the site is collocated with the Cape Grim Baseline Air Pollution Station. CAPE-k ARM aerosol data are available to users and ARM is actively working to ingest these data sets as appropriate. As ARM continues to expand and enhance capabilities, it must scale back in other areas.

4.1.2 Hygroscopicity Plan

4.1.2.1 Operations

For FY25, ARM will operate a single HTDMA at AMF1 (CoURAGE). After the completion of CoURAGE, ARM will retire all HTDMA instruments except one that will be mothballed for potential future use.

Additionally, ARM currently deploys five dual-column CCN instruments. The instruments are operated with one column scanning over pre-defined supersaturation set points while the other column is kept at a constant supersaturation set point for reference. ARM will continue with CCN operations.

The HTDMA instrument currently does provide size-resolved hygroscopicity measurements, including kappa; however, the HTDMA measurements are in the subsaturated regime (RH at 85-90%) while the CCN operates in supersaturated mode.

In limiting the operations of the HTDMA, the goal is to allow users to request different modes of operations as part of limited intensive operational periods (IOPs). Changes to the modes can be requested through the ARM field campaign process for a small field campaign (<https://arm.gov/guidance/campaign-guidelines/small-campaigns>). Potential HTDMA operational modes include:

- Extended humid size scans with higher resolution and/or longer scan times
- More or different dry size cuts (currently 50, 100, 150, 200, 250 nm)
- Ambient size distribution scans (up to 1000 nm) in between regular HTDMA growth factor measurements.

The operational modes listed above are contingent on the outcomes of laboratory tests and may be subject to change. Also, note that some of the listed operational modes are mutually exclusive.

Additionally, ARM currently deploys five dual-column CCN instruments. The instruments are operated with one column scanning over pre-defined supersaturation set points while the other column is kept at a constant supersaturation set point for reference. ARM will continue with CCN operations.

The HTDMA instrument currently does provide size-resolved hygroscopicity measurements, including kappa; however, the HTDMA measurements are in the subsaturated regime (RH at 85-90%) while the CCN operates in supersaturated mode.

4.1.3 Data Products

ARM b-level HTDMA data includes variables for the growth factor and the hygroscopicity parameter kappa (at subsaturated conditions), calculated from the raw instrument data.

As noted in the FY24 Aerosol Operations Plan (Theisen et al 2024), ARM has started to produce a CCN-SMPS kappa product. A CCN and SMPS are deployed at all sites with the exception of NSA. The b1-level HTDMA data products also have a kappa that is based on the simple inversion that the HTDMA is performing in the software. Following the recent retirement of the HTDMAs, the future direction of the kappa VAP remains to be determined.

4.1.4 Ice Nucleating Particles

ARM has established routine ice nucleating particle (INP) filter collection and offline measurements at select sites since 2020. The filters are processed using an ice nucleating spectrometer (INS) at Colorado State University (CSU; also known as the Ice Spectrometer). For FY25, collections of duplicate filters integrated for 24 hours will occur approximately every six days at SGP (C1) and BNF (M1). ARM plans to establish similar measurements at NSA in FY25. AMF collections of duplicate filters every six days currently include the Cloud And Precipitation Experiment at Kennaook (CAPE-k [KGC]) and two sites for CoURAGE (M1 and S2). Daily collections of single filters for 24 hours will occur during a CAPE-k IOP Feb-Apr 2025, and a subset will be processed for case study evaluation (PI: Kerri Pratt). INP filter samples will also be collected on the tethered balloon system (TBS) for six two-week deployments during CoURAGE and at BNF. Data from these sites and campaigns will be available on Data Discovery within 3-6 months after samples are shipped to CSU. We have caught up on the backlog of samples from all previous campaigns and sites.

4.2 Calibration

4.2.1 Calibration Schedule

Generally, the mentors calibrate the aerosol instruments at the beginning and end of mobile facility deployments with calibrations planned in between depending on the length of the deployment and scientific drivers for those campaigns. To align with needs expressed by the community for more routine calibration to better support field campaigns, IOPs, and long-term measurements in general, a new calibration plan for AOS instruments is being developed. The calibration plan is subject to change if other priorities and complications arise. The aerosol team will work with ARM communications on a communication plan to best convey changes to the community. The calibration plan for FY25 is shown in Table 5 with specific instrument-level calibration needs in Table 6. The aerosol mentor team will also

work with site operators to define calibration procedures and transfer ownership of some calibration activities to site operations to improve the timeliness of calibrations and reduce the travel burden on the mentor team.

4.2.2 Intensive Operational Periods

As noted in Table 5, ARM will define periods at SGP, CoURAGE, CAPE-k, and BNF that are ideal for IOP operations in FY25. These periods are determined based on the mentor team's available effort and schedule for the coming years, deconflicted with planned IOPs as much as possible. During these periods, ARM will prioritize efforts to ensure that all instruments are operational, calibrated, and producing high-quality data. This will include shifting known activities (repairs, upgrades, etc.) that would take an instrument offline to either before or after this period and providing more frequent reviews of the data from the Data Quality Office and mentor. These periods do not exclude any other time of the year for IOPs but will be a period in which ARM makes a concerted effort to provide high-quality data with high uptime. In the case of BNF, the IOP periods may include TBS flights.

Table 5. Planned calibration activities for each of the ARM observatories.

Site	Oct 2024	Nov 2024	Dec 2024	Jan 2025	Feb 2025	Mar 2025	Apr 2025	May 2025	Jun 2025	Jul 2025	Aug 2025	Sep 2025	Oct 2025	Nov 2025	Dec 2025
AMF1 (CoURAGE)															
AMF2 (CAPE-K)*															
AMF3 (BNF)															
ENA															
SGP															
NSA															
Calibration Trip															
X – Every effort will be made to ensure															
Ideal periods for IOPS or scheduled IOP															
*CAPE-k operators will be performing these calibrations															

Table 6. Calibration frequency and needs for each instrument type.

Personnel	Instrument	Calibration Frequency	Equipment Needs
Mentor/Operators	ACSM	4x/year	CS, AS, FM, CM
Operators	AETH, flows	1x/year	FM
Mentor/Operators	APS	4x/year	FM
Operators	CAPS	2x/year	AS
Mentor	CCN	2x/year	CS, AS, CM
Operators	CO, MFC	2x/year	FM
Mentor/Operators	CPCf, CPCu/f, SMPS	4x/year	CS, CM
Mentor	HTDMA	2x/year	FM

Personnel	Instrument	Calibration Frequency	Equipment Needs
Mentor	INS/INP, MFC	1/year	FM
Operators	NEPH	2x/year	Cal. gas
Operators	OPC	4x/year	FM
Operators	OZONE, MFC	2x/year	FM
Operators	PSAP, Flow, MFC	2x/year	FM
Operators	SO ₂ , MFC	2x/year	FM
Mentor	SP2	2x/year	CS, AS, CM
Mentor (Operator for calibration check)	UHSAS	2x/year	AS, CM

Cal SMPS (CS)

Atomizer system (AS)

Flow meters (FM)

Calibration material (CM) PSL, ammonium sulfate, nitrate, Fullerene soot

4.2.3 Improving Calibration Processes

In FY24 ARM started the procurement of instrumentation for the establishment of a gold-standard reference for size distribution (SMPS) and number concentration measurements (CPC). This equipment will be installed and tested at the Center for Aerosol Measurement Science (CAMS) at Brookhaven National Laboratory (BNL). These standards will then be compared against World Calibration Center for Aerosol Physics (WCCAP) standards every year. We are seeking to better facilitate these activities between involved organizations (i.e., ARM, CAMS, WCCAP).

Calibrations of ARM CPCs and SMPSs will start in FY25 with closure and intercomparison efforts in FY26/FY27. The goal is, in time, to expand calibrations to other aerosol properties related to the ARM AOS measurements (i.e., composition, optical, hygroscopicity). Table 7 presents the timeline for activities and deliverables in FY25. We expect that calibration activities will result in downtime of the instruments, and they will be planned accordingly as part of the future Aerosol Operations Plan for transparency. The timeline has shifted by about six months compared to the previous AOP due to delays in instrument procurement.

Table 7. CAMS calibration activities for FY25.

Tasks	Deliverables	Planned End Date
Complete first-phase infrastructure preparations	CAMS calibration lab is set up and ready to receive reference instrument for SMPS calibrations.	May 2025
Procurement and installation of reference instrument for SMPS calibrations	Instrument has been installed at BNL and tested for proper operation. Staff has been trained on its operation. Certificate of performance from WCCAP for reference instrument.	July 2025
First calibrations of ARM SMPS instruments using reference instrumentation (starting with SGP), data analysis	ARM SMPSs from SGP have a certificate of performance from CAMS calibration lab.	September 2025
Preparation of lab for CPC calibrations	Calibration lab is set up and ready to receive reference instrument for CPC calibrations.	November 2025
Procurement and installation of reference instrument for CPCs calibrations	Instrument has been installed at BNL and tested for proper operation. Staff has been trained on its operation. Certificate of performance from WCCAP for reference instrument.	January 2026
First calibrations of ARM CPC instruments using gold standards (starting with SGP), data analysis	ARM CPCs from SGP have a certificate of performance from CAMS calibration lab.	February 2026
Documenting and sharing the results and lessons learned from the first CAMS calibrations	Calibration results, procedures, and lessons learned disseminated to relevant parties.	September 2025 (SMPS) and April 2026 (CPC)

4.2.4 Training Site Operations for Calibrations

To reduce travel costs while also maintaining reliable calibrations, the aerosol mentor team has been training site operators in necessary calibrations and maintenance. The first training was for the site operators of the CAPE-k deployment (APS, SP2, UHSAS, SMPS cal system) and was held in November-December 2023. The second training, on April 2024, was for the BNF and SGP site operators. A third training is planned for the AMF1 operators, and it will happen at the CRG deployment. This training will be tailored to the site operators' specific needs. The past training events took place at BNL over the course of one week each time.

4.3 FY25 Engineering and Development

Besides routine AOS operations, the mentor team is focused on improving and expanding ARM measurements to better serve the research community. These efforts include improved understanding of the measurements, new instrumentation, new instrument development, or other needs as they arise. It is important that ARM prioritize these efforts and communicate those priorities to the community. Table 8 shows an overview of activities with further information below for Priority 1 activities.

Table 8. FY25 planned aerosol engineering and development activities.

Task	Priority	Planned End Date
Aerosol node development	1	Dec 2026
Aerosol flux measurement development	1	Dec 2027
Center for Aerosol Measurement Science (CAMS)	1	Dec 2027
Setting up SMPS and CPC gold-standard calibration capabilities for ARM AOS instruments	1	Dec 2026
Absorption Measurement Intercomparison and Calibration Experiment for advancing ARM's absorbing aerosol measurements	1	Sept 2025
Develop and implement a plan for distributed aerosol measurements for urban AMF deployments	1	Dec 2025
Contracted deployment of a NO _x system with the ARM AOS for CoURAGE	1	Dec 2025
Gas Analyzer Intercomparison	1	May 2025
Correction factors for the new PSAP filter media	2	Sept 2025
Refurbishing the ENA HTDMA	2	April 2025 (closed)
Testing a new field-deployable calibration system	2	April 2025 (closed)
Replacement of the aging NEPH systems	2	Dec 2026
Develop a data flagging approach for identifying MOSAiC AOS contamination events	2	April 2025
Evaluating UHSAS and OPC measurements	2	April 2025
Deployment of an OPC inside the CRG AOS container	2	Nov 2025
ACSM Calibration Uncertainty Study	2	April 2025
Drying the AOS impactor and APS sample flows	2	Aug 2025
Retire humidified tandem mobility analyzer (HTDMA)	3	Nov 2025
Modernization of the AOS instrument management software	3	On hold
Gradually remove the real-time flow sensor from the existing standalone CPCs (CPCF, CPCuf) in SGP and ENA in FY24	4	Aug 2025

4.3.1 Aerosol Node Development

As noted earlier, ARM is investing in the development of two aerosol nodes that will initially measure aerosol size distribution and number concentration but may be expanded to include additional measurements such as optical properties, CCN, INP, trace gases, and potentially more.

Aerosol Node Development as Part of the AMF3 Deployment at BNF

Progress update, April 2025

AOS Mini 1 is currently under laboratory development at BNL. The table below shows the adjusted timeline proposed to ARM at the start of FY25 and subsequent progress made on the items listed.

Table 9. Proposed FY25 milestones and progress so far.

Timeframe	Milestone	Activities	Status
FY25	System Integration Continues	Integrate instruments, inlet system and enclosure 1. Inlet system integration with enclosure 2. Test sample conditioning (<40%) and flow control. 3. Characterize/probe particle loss/transmission efficiency 4. Develop and integrate software for instrument/sensors and diagnostics 5. ARM remote connect and data ingest development	Green-completed. Yellow in progress Red- not started.
FY25	System Testing	Two-phase testing to ensure operational stability, field readiness, and reliable continuous measurements Laboratory (controlled) and Ambient (BNL, AMF3). 1. Establish calibration/check procedures for instruments and systems. 2. Comparison with other similar systems (mini-AOS, AOS). 3. Assess lower-cost, research grade portable instruments as a working reference (lab/field) (Metrics for evaluation are still being defined)	

4.3.2 Distributed Aerosol Size Distribution Network for CoURAGE

ARM is also investing into the development of three micro-AOS nodes that measure aerosol size distribution, temperature, and relative humidity. The micro-AOS nodes may be expanded in the future to include additional small sensors.

Progress update, April 2025

Design and engineering review are complete, and three systems are in operation at three ground sites for the CoURAGE campaign in and around Baltimore, Maryland. Between November 2024 and April 2025 all units functioned as expected with good data quality and no data outages. A spare POPS is available to swap failing equipment on short notice.

4.3.3 Aerosol Flux Measurement System (AFS) Development

ARM is deploying a single-particle flux system on the 40-meter tower at AMF3 to determine the viability of measuring aerosol fluxes. If the tests prove successful, additional heights will be deployed in subsequent years.

Progress update, April 2025 (only one milestone for FY24)

Table 10. Revised milestones for the AFS in 2025.

Timeframe	Milestone	Activities	Results
FY25	1. Determine AFS Instrumentation	Survey instruments, procure test units for lab evaluation.	
FY25	Procure AFS instruments and accessories	OPC, data logger, enclosure etc.	
FY26	2. DAQ Development	Maximize AFS sample frequency, characterize time-lag/synchronization.	TBD
FY26	3. System Integration & Testing	Laboratory testing for system stability, time-lag, data flow, and eddy covariance method.	TBD

4.3.4 Center for Aerosol Measurement Science (CAMS) and Setting up SMPS and CPC Gold-Standard Calibration Capabilities for ARM AOS Instruments

As outlined in the Calibration Section (4.2), ARM is enhancing calibration processes for aerosol instruments in the AOS. The initial focus will be on calibrating AOS CPCs and SMPSs using standardized equipment and traceable methods. To support this effort, a calibration plan has been established (see Table 7, Section 4.2.3), new calibration equipment is being procured, and laboratory space at BNL is being prepared. Once the instrument procurements are complete, the CAMS engineering activity will be closed, and these efforts will continue under the activity of setting up SMPS and CPC gold-standard calibration capabilities for ARM AOS instruments.

4.3.5 CARGO-ACT

A dialogue between ARM and ACTRIS, looking for ways to collaborate, started in 2019. What emerged from that dialogue is a three-year project with the European Union named CARGO-ACT – Cooperation and AgReements enhancing Global interOperability for Aerosol, Cloud, and Trace gas research infrastructures. CARGO-ACT’s goal is to deliver a clear roadmap for sustainable global cooperation between key organizations in Europe (ACTRIS) and in the United States (ARM, NOAA, National Aeronautics and Space Administration [NASA]) to provide all users with the best possible services (data, measurement facilities, reference instruments, reference standards, laboratories) for accessing and using information for monitoring climate- and air quality-relevant properties of aerosols, clouds, and trace gases. The project started in March 2024 and ARM contributes in different ways to the six work packages (WP). WP1 deals with data interoperability, WP2 has to do with the specification and documentation of common operation procedures and data quality methodologies, WP3 is the pilot implementations demonstrating service integration, and WPs 4 to 6 deals with governance, common data access, and strategies for how to bring more U.S. observation networks to the ACTRIS project. ARM aerosol operations (in situ and remote sensing) and calibration activities (as part of CAMS) are part of WP2. Deliverables for this year include (1) identification of opportunities for harmonized calibration and operation practices and data production software, (2) agreement on common vocabulary for describing instrument traceability and calibration, quality assurance and quality control, and (3) recommendations for common calibration and operation procedures. The assessment of traceability, calibration, and operational protocols for in situ aerosol measurements (ARM AOS, NOAA Federated Aerosol Network,

and ACTRIS) and the identification of opportunities for harmonization in operation, measurement traceability, and calibration methodology have been completed.

In the context of CARGO-ACT, it is important to highlight that in July 2025, the WMO GAW, through TROPOS WCCAP, will conduct an audit of the SGP AOS. This evaluation, led by Dr. Alfred Wiedensohler, Director of the Central Facility "European Center for Aerosol Calibration & Characterization" within ACTRIS and Head of the World Calibration Center for Aerosol Physics, will assess compliance with established guidelines, provide recommendations for improving measurements, and promote the submission of high-quality data. Such evaluations are crucial for ensuring the accuracy, consistency, and comparability of aerosol measurements, upholding scientific standards, supporting reliable long-term assessments of climate and air quality, and fostering international collaborations. Following this audit, SGP will be designated as a WMO GAW regional station.

4.3.6 Absorption Measurements InterComparison Experiment 2 (AMICE-2)

AMICE-2, to be conducted in summer, 2025, will use in situ instruments (e.g., cavity attenuated phase shift monitor-single-scattering albedo [CAPS-SSA], photothermal interferometer [PTI], photothermal aerosol absorption monitor [PTAAM]) to establish a reference against which to calibrate and correct filter-based instruments recommended for replacement of the PSAP through AMICE-1 (conducted in summer 2024). AMICE-2 will also evaluate the use of prepared solutions to generate internally mixed aerosols with repeatable SSA as references for use in the field. The AMICE final report including recommendations for replacement of the PSAP, and a recommended approach for measurement of absorbing aerosols by ARM, is due by end of FY2025.

4.3.7 Distributed Aerosol Measurement for Urban Deployments

ARM will be deploying three ground-based POPS systems at the main and two supplemental sites for the CoURAGE campaign in Baltimore, Maryland. These are currently under development and plan to be ready for an operational start date of December 1, 2024.

4.3.8 New Trace Gas Analyzer

A new trace gas analyzer (MIRO MGA-10) was evaluated in FY2024 and FY2025 to serve as a replacement for the existing analyzer, along with consolidating the remaining trace gas measurements (O_3 , SO_2) into a single instrument. Regrettably, this new analyzer did not pass acceptance testing and will not be deployed as part of the AOS.

4.3.9 CoURAGE NO_x Measurements

A CAPS-based NO_x system was installed on December 20, 2024 at CoURAGE by the vendor and has been working well with high data availability. A quality-controlled data set will be provided to ARM within six months of the end of the campaign.

5.0 ARM AOS Data Products

5.1 FY24 Accomplishments and Ongoing Activities

Aerosol Optical Depth (AOD) and AOD Best Estimate (AODBE): ARM recently upgraded the MFRSR to include a new 1625-nm channel. This channel has significant interference from gas-phase species and a procedure for subtracting these interferences was developed and implemented in FY23. In FY 24, we processed and released AOD data for the TRACER, Surface Atmosphere Integrated Field Laboratory (SAIL), and Eastern Pacific Cloud Aerosol Precipitation Experiment (EPCAPE) campaigns that includes the new channel (seven channels total). AODBE evaluates the AOD measurements of several co-located instruments and provides a recommended AOD value and error range. We processed and released AODBE data for the ENA and NSA sites.

CCN Kappa: Kappa is a parameterized representation of an aerosol particle's hygroscopicity widely used in models to calculate CCN concentrations and cloud properties. In the last two years, we have developed the CCN-SMPS kappa VAP and processed data for most sites. CCN kappa runs in near-real time and is currently available for all sites/campaigns.

CCN Profile: In previous years, we began implementing a VAP that estimates the CCN vertical profile using lidar data. We recently implemented changes in the VAP that allow it to be run for more sites and time periods. In FY24 we released processed and released data for ENA and additional data for SGP. We also worked to improve the QA/QC implemented in the VAP and have been evaluating the VAP output against available in situ data sets.

Merged Size Distributions: We recently developed a VAP that merges data from the SMPS and APS instruments into a single size distribution and a machine learning algorithm that performs QA/QC on the merged SMPS-APS data set. In FY24, we released merged size distribution data for TRACER and EPCAPE. We also processed merged SMPS-UHSAS size distributions; however, upon investigating the data we noticed a systematic discrepancy between the SMPS and UHSAS particle counts over their common range. We therefore decided to postpone processing and releasing more SMPS-UHSAS merged size distribution data.

Additional VAPs: The ACSM, corrected for composition-dependent collection efficiency (ACSM-CDCE) VAP corrects the ACSM data for non-unity particle sampling. This VAP is now operational and running in real time. Data are available for all sites with an ACSM. The AOP VAP combines PSAP and nephelometer data to provide corrected in situ aerosol optical properties. The AOP VAP also runs in real time and is available for all sites with a PSAP and nephelometer.

5.2 FY25 Planned Activities

Detailed plans for specific data product development efforts in FY25 are as follows.

CCN Profile: In FY25, we plan to continue evaluating the performance of the CCN profile VAP against available in situ data. We will also develop a machine learning-based QA/QC evaluation of the data. Finally, we will evaluate the feasibility of using more values of the CCN supersaturation to retrieve the

CCN vertical profile and explore the possibility of producing an altitude-dependent CCN spectrum across the range of measured saturation values.

HTDMA kappa: In FY24, we began developing a VAP that calculates kappa under subsaturated conditions using data from ARM’s HTDMA instruments. This kappa calculation is less sensitive to errors in particle count than the CCN kappa, only relies on data from a single instrument, and provides a measure of kappa under subsaturated conditions. In FY25 we will finalize the development of this algorithm and produce HTDMA kappa data and compare data to the CCNC-SMPS-based kappa.

Merged Size Distribution Machine Learning: In FY24, we developed a machine learning (ML) algorithm that performs QA/QC labeling of the merged SMPS/APS size distribution data. We will continue to evaluate the performance of this algorithm with data from new sites and release ML-labeled merged-size distribution data.

AOD and AODBE: We have been working on extending the AOD VAP to additional sites that require more manual labor in evaluating the NASA Langley Research Center measurements and will continue to do so in FY25, including processing AOD data for new campaigns such as the ARM West Antarctic Radiation Experiment (AWARE) and the BNF site. In FY25 we will incorporate data from the new 1625 channel into the AODBE VAP, creating a seven-channel product, and process data for TRACER, SAIL, and ECAPE.

Bundled Aerosol Data: At the request of ARM users and advisory groups, we will develop a VAP that bundles data from multiple aerosol instrumentation into a single file on a single time stamp. This VAP will simplify the user experience and reduce the amount of work for scientists wanting to use data from multiple instruments in their analysis.

6.0 References

- Aiken, AC, F Gallo, J Uin, SR Springston, J Wang, G Zheng, C Kuang, A McComiskey, R Wood, C Flynn, A Theisen, EB Azevedo, P Ortega, and H Powers. 2019. “Eastern North Atlantic (ENA) Aerosol Supplementary Site (S1) Data Analysis Report.” U.S. Department of Energy. DOE/SC-ARM-TR-229. <https://doi.org/10.2172/1567063>
- Gallo, F, J Uin, S Springston, J Wang, G Zheng, C Kuang, R Wood, EB Azevedo, A McComiskey, F Mei, A Theisen, J Kyrouac, and AC Aiken. 2020. “Identifying a regional aerosol baseline in the eastern North Atlantic using collocated measurements and a mathematical algorithm to mask high-submicron-number-concentration aerosol events.” *Atmospheric Chemistry and Physics* 20(12): 7553–7573. <https://doi.org/10.5194/acp-20-7553-2020>
- Mather, J, S Springston, and C Flynn. 2018. ARM Aerosol Measurement Plan. U.S. Department of Energy. DOE/SC-ARM-TR-213. <https://doi.org/10.2172/1437623>
- Middlebrook, AM, R Bahreini, JL Jimenez, and MR Canagaratna. 2012. “Evaluation of composition-dependent collection efficiencies for the aerodyne aerosol mass spectrometer using field data.” *Aerosol Science and Technology* 46(3): 258–271, <https://doi.org/10.1080/02786826.2011.620041>

Theisen, A, O Mayol-Bracero, J Uin, S Smith, J Shilling, C Kuang, M Zawadowicz, A Singh, R Trojanowski, J Creamean, A Sedlacek, C Hayes, D Campos DeOliveira, and M Allain. 2023. ARM FY2024 Aerosol Operations Plan. ARM user facility. DOE/SC-ARM-TR-289, <https://doi.org/10.2172/2008425>

Uin, J, AC Aiken, MK Dubey, C Kuang, M Pekour, C Salwen, AJ Sedlacek, G Senum, S Smith, J Wang, TB Watson, and SR Springston. 2019. “Atmospheric Radiation Measurement (ARM) Aerosol Observing Systems (AOS) for Surface-Based In Situ Atmospheric Aerosol and Trace Gas Measurements.” *Journal of Atmospheric and Oceanic Technology* 36(12): 2429–2447, <https://doi.org/10.1175/JTECH-D-19-0077.1>

Appendix A

2024 Aerosol Operations Plan Completed Tasks

- **Improve and simplify access to aerosol data**

We have made substantial improvements in aerosol data access and are also continually improving based on user feedback. Key improvements:

- New Data Discovery interface
- Additional metadata to improve the search capabilities
- Spatial, temporal, and keyword search capabilities
- Dedicated home page search by Category for “Aerosols”
- Improved data details page with contact information, citations, data quality timeline, primary measurements, and data plots
- Added recommendations for aerosol datastreams.

- **Improve documentation of measurements and datastreams**

Metadata auditing for the aerosol datastreams is complete. The ARM Data Center metadata team improved the metadata keywords and classifications.

The VAP web pages have been updated to provide improved short summary descriptions of VAPs including caveats that describe when the VAP is applicable and when caution should be used. The reference list has been updated and made easier to use to identify VAP point of contacts.

CARGO-ACT work package 2 will also contribute to improve documentation regarding aerosol measurements (in situ and remote sensing).

- **Reduce number of HTDMAs in field**

ARM has reduced the number of operating HTDMAs to two at a maximum. For FY25, this will include BNF and AMF1 (CoURAGE). Once CoURAGE is completed, the HTDMA at SGP will be brought online.

- **Refurbishing of ENA HTDMA**

This is not happening, since after the completion of CoURAGE ARM is retiring all HTDMA instruments.

- **Removal of ACSM at ENA**

In FY24, ARM retired from operations the ENA ACSM to add an ACSM to NSA.

- **Retirement of humidigraphs**

To enable new scope, in FY24 ARM retired all humidigraphs and RH-scanned nephelometers from operations.

- **New instrumentation for AMF3**

The AMF3 will be deployed at BNF in Alabama, starting September 2024. In addition to the aerosol node and aerosol flux measurement development, ARM added additional instrumentation to the AMF3 AOS. This includes an APS, SO₂, extended-size-range SMPS (size range is extended from 10-500 nm to 10-800 nm), and SP2 (SP2-XR). The SO₂ and APS are standard ARM instruments, but the SMPS and the SP2-XR are new additions that require ARM to develop new processes (ingest, data quality, calibration, etc.).

- **NSA additional instrument deployment**

In September 2024, ARM started deployment of the following instruments: ACSM, APS, SP2-XR, and filters for INS at the NOAA facility alongside NOAA instrumentation.

- **Develop prioritization for ultrafine CPCs**

We have “increased” the number density of the ultrafine CPC (CPCu or CPCuf) to four ARM sites (AMF1, AMF2, AMF3, SGP).

- **Implement second set of size distribution instruments**

There are SMPSs and APSs in all five ARM deployments. The new SMPS at the AMF3 has the wide-range differential mobility analyzer that allows a measurement range of 10 to 800 nm. There is a sixth APS at NSA since September 2024.

- **Comparable size distribution output across instruments**

The aerosol translator is leading an effort to create a merged size distribution VAP using a machine learning method. This is under testing and evaluation.

- **Absorption, scattering, and extinction closure**

This was an optical closure analysis involving the Aerodyne CAPS (cavity attenuated phase shift spectroscopy) aerosol extinction monitor and the TSI nephelometer (aerosol scattering). The particle soot absorption photometer (PSAP, aerosol absorption) was to be the third element in this study, but due to the discontinuation of the filter media, it was not included. The study demonstrated that measurements of scattering coefficients and extinction coefficients for purely scattering particles agree to within 5-10 % in controlled laboratory experiments. This accuracy limits the extent to which closure between scattering, absorption, and extinction can be attained for an aerosol that contains absorbing components. The fact that numerous correction schemes have been, and are still being, proposed for filter-based measurements of aerosol light absorption (e.g., PSAP or aethalometer), which can vary considerably among themselves, indicates that this approach is unlikely to yield accurate results under any situation. For this reason, filter-based measurements are really only useful in identifying relative changes in absorption, not the actual absorption coefficients.

- **Increase PSAP filter change at SGP**

On a nightly basis, every weekend, and during extended weekends, the PSAP filters at SGP face an excessive load, resulting in the loss of absorption measurements for several hours to an entire day each weekend, contingent on local conditions. To extend the operational window of the SGP PSAP, we increased the filter change frequency at minimally two changes per day—at the start and end of each shift.

- **New trace gas analyzer**

We tested the MIRO Gas analyzer (MGA-10) as a potential replacement for ARM trace gas instruments and it is not a viable option for ARM field deployments. See Section 4.3.7.

- **Develop plan to support detailed composition measurements**

ARM's plan is to operate the ACSMs and look at opportunities for expansion as they present themselves and as budgets align.

- **Develop strategy for hygroscopic measurements**

Following the reduction in the HTDMA operations, the freed-up HTDMA units are undergoing performance evaluation before starting the tests on new operational modes. For more details regarding this plan please refer to Section 4.1.2.

- **Upgrade an ARM HSRL to support aerosol retrievals**

ARM upgraded one high-spectral-resolution lidar (HSRL) to add a 532-nm wide-field-of-view channel and 1064-nm narrow-field-of-view channel along with off-nadir scanning. This system is deployed at SGP, alongside the Raman lidar (RL) which will allow for three-wavelength retrievals of aerosols optical and physical properties. ARM is in the process of upgrading its second HSRL to the same configuration for deployment at the AMF3 observatory in Bankhead National Forest where there will also be a RL deployed. ARM is also in the early phases of procuring an aerosol profiling lidar, similar to those deployed for ACTRIS. The plan is to incorporate aerosol microphysical retrieval into a VAP. ARM is working with the NASA Langley Research Center to receive their TIARA algorithm (Tikhonov Advanced Regularization Algorithm), which uses HSRL and RL to derive aerosol properties. In the last year, data analyst Peng Wu has developed an aerosol feature mask that identifies aerosol features using HSRL data.

Appendix B

2024 Aerosol Measurement In-Progress Tasks

- **Identify candidate data products for other communities**

ARM has been actively working toward making its data available through the Global Atmosphere Watch (GAW) database, and those efforts are ongoing. In parallel, ARM has engaged with ACTRIS to align on common practices and data products. These collaborations led to the launch of the three-year CARGO-ACT project (Cooperation and Agreements enhancing Global interoperability for Aerosol, Cloud and Trace gas research infrastructures). The aim of CARGO-ACT is to harmonize aerosol data across ARM, ACTRIS, NOAA, and NASA, thereby improving usability for the global research community, including modelers and satellite data users.

As part of this interoperability initiative, the ARM Data Center provided a list of candidate data sets and a FAIR (Findability, Accessibility, Interoperability, and Reusability) implementation plan to CARGO-ACT.

Additionally, ARM has successfully harvested metadata records from ACTRIS and made them discoverable through the ARM Data Discovery tool. Users can now search and explore these external data sets via ARM's interface, while downloads are seamlessly redirected to the ACTRIS portal.
https://adc.arm.gov/discovery/#/results/data_repository::ACTRIS

- **Refurbishing of ENA HTDMA**

The ENA HTDMA is being refurbished and will be one of the HTDMA units that are undergoing performance evaluation before starting the tests on new operational modes.

- **Implement an inlet drying system (initially at SGP)**

This activity was to track the development and testing of a drying system for the AOS impactor line and the APS at SGP. With the pending installation of a Nafion dryer for the SGP APS (planned for our next SGP trip), this initial scope of work will be complete.

- **Implement new PTRH sensors in AOS02 and AOS03**

A new PTRH sensor was developed and successfully deployed in AOS01 and AOS03. These sensors are in each sample line, as close to the instrument as possible to characterize the conditions of the sample air as it enters the instruments. It will be installed in AOS02 before its next deployment.

- **Support lidar/radiometer retrieval development**

A translator is currently preparing the Raman Lidar Vertical Profiles-Feature Detection and Extinction (RLPROF-FEX) VAP at SGP and improving the HSRL data to support multi-wavelength lidar retrievals of aerosols.

The development of the aerosol microphysical properties retrieval has experienced some delays. We are currently awaiting the code from NASA, as they are in the process of finalizing the open-source code release. In the meantime, we have been focusing on formalizing the processing workflow and collaborating with NASA to verify that we have the correct input data. NASA is actively testing our input data using their algorithms to ensure that we can successfully reproduce the results from the Cryogenic, High-Accuracy Refraction Measuring System (CHARMS) campaign.

Appendix C

Download Metrics 1/1/2018-7/7/2023

Does not include aircraft data.

Table 11. Instrument download metrics, 2018-2023.

Instrument	Non-Infrastructure Downloads	Non-Infrastructure Users	Publications
CCN*	1072	333	15
CPC	772	259	20
SMPS*	771	230	23
AOSMET	514	205	4
UHSAS	519	190	7
NEPH	487	163	13
ACSM*	422	145	8
PSAP	398	135	9
AOS	395	180	30
CO	316	93	2
HTDMA	264	83	4
OZONE	177	71	4
SP2*	172	83	12
AETH	123	67	4
APS*	107	46	6
SO2	79	39	1
GHG	64	16	2
CAPS-PMEX	58	20	2
OPC	43	12	2
NOX	42	25	0
CLAP	39	19	1

*Numbers include VAPs, if available, which may pull in data from multiple products.



www.arm.gov

U.S. DEPARTMENT OF
ENERGY

Office of Science